

APPENDIX B
VERSION WITH MARKINGS TO SHOW CHANGES MADE
37 C.F.R. § 1.121(b)(iii) AND (c)(ii)

SPECIFICATION:

Paragraph at page 7, line 28, to page 8, line 5:

To further reduce the size of the emitter [is mounted] , the emitter may be mounted on a "THz window". The window material could be for example polyethylene, polythene, high-resistivity silicon, Z-quartz or TPX (poly-4-methylpentene-1), it must be at least substantially transparent to the irradiating beam. The window would preferably be thin, for example, between 50 and 300 microns. This is to ensure that the THz beam diameter is still smaller than the shortest wavelength component when the THz beam reaches the sample. The size of the window is large enough to allow all of the reflected beams to be collected with negligible loss. As the emitter is provided on the window which is substantially transparent to THz, the THz can pass through the mount for the emitter.